

Made 06/30/23

Page: 1

Date: 06/30/23 at 3:44 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Package	a i ioaaoti	54		
	SODIUM STANDARD 1.0 PPM N.I.S.T. TF			
Test		Target/UOM	Range	Result
	D TO FORMULATION			YES
FILE		YES		
	Sample ID			
	Tested 06/30/	23		
	Tester AROS	}		
	Recorded 06/30/	23		
PPM SODI	UM	1.00	0.98 - 1.02	1.00
		PPM		
	Sample ID			
	Tested 06/30/	23		
	Tester AROS			
	Recorded 06/30/			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
	Sample ID			
	Tested 06/30/	23		
	Tester AROS			
	Recorded 06/30/			
INSTRUME	ENTS USED DURING			P-651,VF3-J3
PREPARA	TION			,
	Sample ID			
	Tested 06/30/	/23		
	Tester AROS			
	Recorded 06/30/			
SHELF LIFE OF PRODUCT		6 MONTHS		6 MONTHS
	. 2	TIME		
	Sample ID			
	Tested 06/30/	23		
	Tester AROS			
	Recorded 06/30/			
EXPIRATION		MM/DD/YYYY		01/30/24
		MM/DD/YYYY		0.,00,=.
	Sample ID	, 22,		
	Tested 06/30/	93		
	Tester AROS			
	Recorded 06/30/			
	t# 3063091	20		



Page: 2

Date: 06/30/23 at 3:44 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

Quality Manager