



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7330			
POTASSIUM IODIDE-IODATE 0.025 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0250 EQ/L	0.0245 - 0.0255	0.0246
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
QCSP33:POTASSIUM IODATE-IODIDE TITRANT	YES QCSP33		YES
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,M112,VF5-L12,M55,BID04 099
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
Actual Density	1.00577 GM/ML	1.00554 - 1.00599	1.00573
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		07/30/24
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		1 YEAR
Sample ID			
Tested	06/30/23		
Tester	MHUS		
Recorded	06/30/23		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

N.I.S.T. Traceable to SRM 136

YES

YES

YES

Sample ID

Tested 06/30/23

Tester MHUS

Recorded 06/30/23

Lot # 3062820

Made 06/30/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager