



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> ERL198F			
SODIUM NITROPRUSSIDE SOLUTION SMART.NH4 TEST			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 06/28/23			
Tester AFRE			
Recorded 06/28/23			
<b>CALCULATED W/V% SODIUM NITROPRUSSIDE</b>	0.050	0.049 - 0.051	0.050
Sample ID	W/V%		
Tested 06/28/23			
Tester AFRE			
Recorded 06/28/23			
<b>APPEARANCE AND COLOR</b>	PALE TAN SOLUTION APPEARANCE		PASS
Sample ID			
Tested 06/28/23			
Tester AFRE			
Recorded 06/28/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M108VF500-X41
Sample ID			
Tested 06/28/23			
Tester AFRE			
Recorded 06/28/23			
<b>EXPIRATION DATE</b>	MM/DD/YYYY		08/02/23
	MM/DD/YYYY		
Sample ID			
Tested 06/28/23			
Tester AFRE			
Recorded 06/28/23			
<b>Lot #</b>	3062615		
<b>Made</b>	06/28/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.