



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: DC662F			
	HYDROCHLORIC ACID 0.01 NORMAL (MR/THAM)		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		06/30/23
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	22.2
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		06/30/23
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
Actual Normality	0.0100	0.0095 - 0.0105	0.0099
	EQ/L		
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
N.I.S.T. Traceable to SRM 723 THAM	YES		YES
	YES		
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 06/30/23			
Tester JENC			
Recorded 06/30/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-204,BID-02253,PID-M91
Sample ID		
Tested 06/30/23		
Tester JENC		
Recorded 06/30/23		
Actual Density	0.99842	0.99831 - 0.99853
	GM/ML	0.99840
Sample ID		
Tested 06/30/23		
Tester JENC		
Recorded 06/30/23		
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0
	DEG C	22.2
Sample ID		
Tested 06/30/23		
Tester JENC		
Recorded 06/30/23		
Lot # 3062830		
Made 06/30/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager