



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LEI012 CYCLOHEXANONE/MEK 70:30 V/V (4L)			
Lot # 3061401			
Chemical Product: LEI012F CYCLOHEXANONE/METHYL ETHYL KETONE 70/30% V/V			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR	CLEAR COLORLESS LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/19/23		
Tester	LWIL		
Recorded	06/19/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	06/19/23		
Tester	LWIL		
Recorded	06/19/23		
GC ANALYSIS	METHYL ETHYL KETONE 28-32% V/V GC ANALYSIS		29.1
Sample ID			
Tested	06/27/23		
Tester	CSAN		
Recorded	06/27/23		
GC ANALYSIS	CYCLOHEXANONE 68-72% V/V GC ANALYSIS		70.9
Sample ID			
Tested	06/27/23		
Tester	CSAN		
Recorded	06/27/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		06/30/26
Sample ID			
Tested	06/19/23		
Tester	LWIL		
Recorded	06/19/23		
Made	06/19/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/27/23 at 8:24 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager