



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> LY396F			
ORGANIC ACID CALIBRATION STOCK 1,000 PPM EA			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>CALCULATED PPM PROPIONATE</b>	1000.00 PPM	995.00 - 1005.00	998.84
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>CALCULATED PPM ACETATE</b>	1000.00 PPM	995.00 - 1005.00	1000.56
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>CALCULATED PPM FORMATE</b>	1000.00 PPM	995.00 - 1005.00	1000.35
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>CALCULATED PPM BUTYRATE</b>	1000.00 PPM	995.00 - 1005.00	1000.40
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>CALCULATED PPM CARBONATE</b>	1000.00 PPM	995.00 - 1005.00	1000.27
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>CALCULATED PPM NITRATE</b>	1000.00 PPM	995.00 - 1005.00	999.68
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/05/23			
Tester AROS			
Recorded 07/05/23			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 07/05/23 at 11:21 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**INSTRUMENTS USED DURING  
PREPARATION**

INSTRUMENT

M108

Sample ID

Tested 07/05/23

Tester AROS

Recorded 07/05/23

**Lot #** 3062097

**Made** 07/05/23

**Quality Manager**