

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Pro	oduct: HON176F			
	Fluoride	Standard 7.0 ppm w/T	isab II	
Test		Target/UOM	Range	Result
	FORMULATION ON	YES		YES
FILE		YES		
San	nple ID			
Tes	ted 07/07/23			
Tes	ter DTRA			
Rec	orded 07/07/23			
QCSP45:MEASURING FLUORIDE IN		YES		YES
WATER		QCSP45		
San	nple ID			
Tes	sted 07/07/23			
Tes	ter DTRA			
Rec	orded 07/07/23			
CALCULATED PPM FLUORIDE		7.00	6.70 - 7.30	7.15
		PPM (MG/L)		
San	nple ID			
Tes	sted 07/07/23			
Tes	ter DTRA			
Rec	orded 07/07/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
San	nple ID			
Tes	ted 07/07/23			
Tes	ter DTRA			
Rec	corded 07/07/23			
INSTRUMENTS USED DURING		INSTRUMENT		GC,M-79
PREPARATION				
San	nple ID			
Tes	••••=•			
Tes				
	corded 07/07/23			
Lot #	3070677			
	07/07/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA