

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8654			
SODIUM HYDROXIDE 1.5 NORMAL SOLUTION			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		07/14/23
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	22.2
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		07/14/23
	MM/DD/YYYY		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	22.2
<b>DEG C</b>	DEG C		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>NORMALITY TO 4 DEC PLACES</b>	1.5000	1.4991 - 1.5009	1.5001
	EQUIV/LITER		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>QCSP7: STANDARDIZATION OF</b>	YES		YES
<b>AQUEOUS BASES</b>	METHOD		
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>N.I.S.T. Traceable to SRM 84</b>		YES	YES
		YES	
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>APPEARANCE AND COLOR</b>		CLEAR LIQUID	PASS
		APPEARANCE	
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>		INSTRUMENT	M103,T-22BID-3902,PID-M91
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>Actual Density</b>		1.06106	1.06075 - 1.06136
		GM/ML	1.06112
Sample ID			
Tested 07/14/23			
Tester MHUS			
Recorded 07/14/23			
<b>Lot #</b> 3071307			
<b>Made</b> 07/14/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**