



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> HON142F Fluoride/Citrate Standard 3.0 ppm Solution			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested 07/12/23			
Tester RIRO			
Recorded 07/12/23			
<b>CALCULATED PPM FLUORIDE</b>	3.00 PPM (MG/L)	2.80 - 3.20	2.90
Sample ID			
Tested 07/12/23			
Tester RIRO			
Recorded 07/12/23			
<b>APPEARANCE AND COLOR</b>	CLEAR PALE BLUE APPEARANCE		PASS
Sample ID			
Tested 07/12/23			
Tester RIRO			
Recorded 07/12/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P64,VF1-Z16
Sample ID			
Tested 07/12/23			
Tester RIRO			
Recorded 07/12/23			
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY		01/30/24
Sample ID			
Tested 07/12/23			
Tester RIRO			
Recorded 07/12/23			
<b>Lot #</b>	30710126		
<b>Made</b>	07/12/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.