



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4370			
HYDROCHLORIC ACID 6.0 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		07/11/23
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	30.7
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/11/23
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	23.0
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
NORMALITY	6.000	5.950 - 6.050	5.971
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			
Actual Density	1.09706	1.09574 - 1.09839	1.09752
Sample ID			
Tested 07/11/23			
Tester CCOA			
Recorded 07/11/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 723 THAM

YES YES
 YES

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

APPEARANCE AND COLOR

CLEAR LIQUID APPEARANCE PASS

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

INSTRUMENTS USED DURING PREPARATION

INSTRUMENT M55,M24,T24

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

EXPIRATION DATE

MM/DD/YY 07/30/26
 MM/DD/YYYY

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

APHA COLOR

5 0 - 10 10
 COLOR UNITS

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

EXPIRATION DATE

MM/DD/YY 07/30/26
 MM/DD/YYYY

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

SHELF LIFE OF PRODUCT

SHELF LIFE TIME 3 YEARS

Sample ID
 Tested 07/11/23
 Tester CCOA
 Recorded 07/11/23

Lot # 3070745

Made 07/11/23



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 07/12/23 at 6:58 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager