

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ERL113F POTASSIUM HYDROXIDE 0.5 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		07/13/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	22.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/13/23
	MM/DD/YYYY		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.0
DEG C	DEG C		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
LOT NUMBER OF ORIGINAL	LOT #		3012430
MATERIAL			
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			
Actual Normality	0.5000	0.4995 - 0.5005	0.5004
	EQ/L		
Sample ID			
Tested 07/12/23			
Tester MHUS			
Recorded 07/13/23			


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Actual Density	1.02322 GM/ML	1.02299 - 1.02345	1.02325
Sample ID			
Tested	07/12/23		
Tester	MHUS		
Recorded	07/13/23		
N.I.S.T. Traceable to SRM 84 (KHP)	YES YES		YES
Sample ID			
Tested	07/12/23		
Tester	MHUS		
Recorded	07/13/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	07/12/23		
Tester	MHUS		
Recorded	07/13/23		
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		M55,BID-3902,PID-M91
Sample ID			
Tested	07/12/23		
Tester	MHUS		
Recorded	07/13/23		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		07/30/25
Sample ID			
Tested	07/12/23		
Tester	MHUS		
Recorded	07/13/23		
Lot #	3071157		
Made	07/13/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager