

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produc	t: SPX2841	F		
	CERIC	SULFATE 0.2 NORMAL IN 2	NORMAL SULFURIC ACID	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES	-	YES
FILE		YES		
Sample I)			
Tested	07/11/23			
Tester	MHUS			
Recorded	07/11/23			
Actual Normality		0.2000	0.1985 - 0.2015	0.2001
		EQ/L		
Sample I)			
Tested	07/11/23			
Tester	MHUS			
Recorded				
Actual Density		1.13095	1.12805 - 1.13385	1.13228
		GM/ML		
Sample II)			
Tested	07/11/23			
Tester	MHUS			
Recorded				
QCSP19:STANDARDIZ		YES		YES
CERIC SULFATE		QCSP19		120
Sample IE	`			
Tested				
	07/11/23			
Tester	MHUS			
		CLEAR ORANGE LIQUID		PASS
APPEARANCE AND C	OLOR			PASS
		APPEARANCE		
Sample II				
Tested	07/11/23			
Tester	MHUS			
Recorded				
INSTRUMENTS USED	DURING	INSTRUMENT		M113,VF3-J21,T-4,M55,BID390
PREPARATION				2
Sample I)			
Tested	07/11/23			
Tester	MHUS			
Recorded	07/11/23			
EXPIRATION DATE		MM/DD/YY		01/30/24
		MM/DD/YYYY		
Sample I)			
Tested	07/11/23			
Tester	MHUS			
Recorded	07/11/23			
Lot # 3071	089			
Made 07/1				
	1/20			



Page: 2 Date: 07/11/23 at 4:33 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard SE

Quality Manager