

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ	u ct: 2160			
_	COD STAN	NDARD 50 PPM W/W AQUEOU	S	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample	e ID			
Tested	07/11/23			
Tester	AROS			
Record	ed 07/11/23			
PPM COD		50.0	47.0 - 53.0	50.8
		PPM		
Sample	e ID			
Tested	07/11/23			
Tester	AROS			
Record				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample	e ID			
Tested	07/11/23			
Tester	AROS			
Record				
INSTRUMENTS USED DURING		INSTRUMENT		P46,VF1-Z31
PREPARATION				
Sample	e ID			
Tested	07/11/23			
Tester	AROS			
Record				
EXPIRATION DATE		MM/DD/YY		09/16/23
-		MM/DD/YYYY		
Sample	: ID			
Tested	07/11/23			
Tester	AROS			
Record				
SHELF LIFE OF PRODUCT		2 MONTHS		2 MONTHS
	-	TIME		
Sample	e ID			
Tested 07/11/23				
Tester	AROS			
Record				
Lot # 30	70734			
	/11/23			
	/ / _ J			



Page: 2 Date: 07/12/23 at 7:06 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard SE

Quality Manager