

Page: 1

Date: 07/12/23 at 7:02 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	9577			
		ANDARD 5000 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	07/11/23			
Tester	BPHI			
Recorded	07/11/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	07/11/23			
Tester	DTRA			
Recorded	07/11/23			
INSTRUMENTS USED		INSTRUMENT		M91,T11
PREPARATION				- ,
Sample ID				
Tested	07/11/23			
Tester	DTRA			
Recorded	07/11/23			
EXPIRATION DATE	07/11/25	MM/DD/YY		01/30/24
EXI IIIANON DATE		MM/DD/YYYY		01/00/21
Sample ID		WWW, BB, TTTT		
Tested	07/11/23			
Tester	DTRA			
Recorded	07/11/23			
		SHELF LIFE		6 MONTHS
SHELF LIFE OF PRODUCT		TIME		0 MONTHS
0		TIIVIE		
Sample ID				
Tested	07/11/23			
Tester	DTRA			
Recorded PPM TOC NO DEC PLA	07/11/23	5000	4750 - 5250	4901
PPM TOC NO DEC PLA	ACES	PPM	4750 - 5250	4901
Sample ID				
Tested	07/11/23			
Tester	DTRA			
Recorded	07/11/23			
Lot # 30710105				
Made 07/11/23				
· -				



Page: 2

Date: 07/12/23 at 7:02 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager