

Page: 1

Date: 07/20/23 at 4:53 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 9096			
. uonagoa i rouaon	C ACID 4.0 NORMAL		
		D	B !!
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/20/23			
Tester DTRA			
Recorded 07/20/23	VEO		VEO
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 07/20/23			
Tester DTRA			
Recorded 07/20/23 DATE OF SAMPLING FOR TEST	MM/DD/VVVV		07/00/02
	MM/DD/YYYY		07/20/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 07/20/23			
Tester DTRA Recorded 07/20/23			
	25.0	10.0 - 35.0	22.5
AMBIENT TEMPERATURE DURING	DEG C	10.0 - 33.0	22.0
SAMPLING DEGC	DLG C		
Sample ID Tested 07/20/23			
Tested 07/20/23 Tester DTRA			
Recorded 07/20/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/20/23
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/20/20
Sample ID	IVIIVI/DD/1111		
Tested 07/20/23 Tester DTRA			
Recorded 07/20/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.5
DEG C	DEG C	13.0 30.0	22.5
	DEG 0		
Sample ID Tested 07/20/23			
Tester DTRA			
Recorded 07/20/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
ALL LABARCE AND COLOR	LIQUID		1 700
	APPEARANCE		
Sample ID			
Tested 07/20/23			
Tester DTRA			
Recorded 07/20/23			
QCSP5: STANDARDIZATION OF	YES		YES
AQUEOUS ACIDS	METHOD		
Sample ID			
Tested 07/20/23			
Tester DTRA			
Recorded 07/20/23			
ı			

07/20/23

Made

Page: 2

Date: 07/20/23 at 4:53 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

NORMALITY TO 4 DEC	PLACES	4.0000	3.9950 - 4.0050	3.9993
		EQUIV/LITER		
Sample ID				
Tested	07/20/23			
Tester	DTRA			
Recorded	07/20/23			
Actual Density		1.12038	1.11955 - 1.12122	1.12089
		GM/ML		
Sample ID				
Tested	07/20/23			
Tester	DTRA			
Recorded	07/20/23			
INSTRUMENTS USED D	URING	INSTRUMENT		VF6-W2,GC,T-38,M55,BID0078
PREPARATION				7
Sample ID				
Tested	07/20/23			
Tester	DTRA			
Recorded	07/20/23			
N.I.S.T. Traceable to SRM 723 THAM		YES		YES
		YES		
Sample ID				
Tested	07/20/23			
Tester	DTRA			
Recorded	07/20/23			
EXPIRATION DATE		MM/DD/YY		07/30/25
		MM/DD/YYYY		
Sample ID				
Tested	07/20/23			
Tester	DTRA			
Recorded	07/20/23			
Lot # 30719	54			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager