

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: 694	15		
CONDUCTIVITY STANDARD 50,000 uMHOS			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION	ON YES	-	YES
FILE	YES		
Sample ID			
Tested 07/17/2	3		
Tester JRIO			
Recorded 07/17/2			
CONDUCTIVITY @25 C.	50000	49999 - 50001	50000
	US/CM		
Sample ID			
Tested 07/17/2	3		
Tester JRIO			
Recorded 07/17/2			
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 07/17/2	3		
Tester JRIO			
Recorded 07/17/2			
TRACEABLE TO N.I.S.T. SRM 37			YES
	YES		
Sample ID			
Tested 07/17/2	3		
Tester JRIO			
Recorded 07/17/2			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 07/17/2	3		
Tester JRIO			
Recorded 07/17/2			
INSTRUMENTS USED DURING	INSTRUMENT		M92,T229
PREPARATION			
Sample ID			
Tested 07/17/2	3		
Tester JRIO	_		
Recorded 07/17/2			07/00/04
EXPIRATION DATE	MM/DD/YY		07/30/24
	MM/DD/YYYY		
Sample ID			
Tested 07/17/2	3		
Tester JRIO			
Recorded 07/17/2			
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
	TIME		
Sample ID	_		
Tested 07/17/2	3		
Tester JRIO	2		
Recorded 07/17/2	3		



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	3071774
Made	07/17/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager