

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPE509F			
TOTAL SUSPENDED SOLIDS STANDARD 1000 PPM			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
APPEARANCE AND COLOR	TURBID LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
CALCULATED MG/L (PPM) TSS	1000.0 MG/L	995.0 - 1005.0	1000.5
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
ACT GMS CELITE	1.0000 GM	0.9990 - 1.0010	1.0005
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
QCSP11:STANDARDIZATION OF BROMIDE-BROMAT	YES QCSP11		YES
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
MLS OF WATER	1000 ML	995 - 1005	1000
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF1-Z2
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			
WEIGHT NIST TRACEABLE	PASS YES		PASS
Sample ID			
Tested 07/17/23			
Tester AROS			
Recorded 07/17/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/17/23 at 4:59 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YY
MM/DD/YYYY

07/30/24

Sample ID

Tested 07/17/23

Tester AROS

Recorded 07/17/23

SHELF LIFE OF PRODUCT

SHELF LIFE
TIME

1 YEAR

Sample ID

Tested 07/17/23

Tester AROS

Recorded 07/17/23

Lot # 3071747

Made 07/17/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager