

Recorded 07/18/23

Page: 1

Date: 07/18/23 at 2:16 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

| Packaged Product: 6466 | | | |
|--|----------------|-----------------|----------|
| NITRIC A | CID 0.3 NORMAL | | |
| Test | Target/UOM | Range | Result |
| Actual Normality | 0.3000 | 0.2995 - 0.3005 | 0.2997 |
| | EQ/L | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |
| Recorded 07/18/23 | | | |
| QCSP5: STANDARDIZATION OF | YES | | YES |
| AQUEOUS ACIDS | METHOD | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |
| Recorded 07/18/23 | | | |
| N.I.S.T. Traceable to SRM 723 THAM | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |
| Recorded 07/18/23 | | | |
| PREPARED TO FORMULATION ON | YES | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |
| Recorded 07/18/23 | | | |
| SAMPLE COLLECTION & HANDLING | YES | | YES |
| PER LAB040 | YES | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |
| Recorded 07/18/23 | MANDDAGGA | | 07/40/00 |
| DATE OF SAMPLING FOR TEST | MM/DD/YYYY | | 07/18/23 |
| PURPOSES | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |
| Recorded 07/18/23 | 25.0 | 10.0 25.0 | 22.5 |
| AMBIENT TEMPERATURE DURING | | 10.0 - 35.0 | 22.5 |
| SAMPLING DEGC | DEG C | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA Recorded 07/18/23 | | | |
| Recorded 07/18/23 DATE OF TESTING OF SAMPLE | MM/DD/YYYY | | 07/18/23 |
| DATE OF TESTING OF SAMPLE | | | U1110/23 |
| Commis ID | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 07/18/23 | | | |
| Tester DTRA | | | |

07/18/23

Made

Page: 2

Date: 07/18/23 at 2:16 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

AMBIENT TEST TEMPERATURE 20.0 15.0 - 30.0 22.5 DEG C DEG C Sample ID Tested 07/18/23 Tester **DTRA** Recorded 07/18/23 APPEARANCE AND COLOR **CLEAR WATER WHITE PASS** LIQUID **APPEARANCE** Sample ID Tested 07/18/23 Tester **DTRA** Recorded 07/18/23 INSTRUMENTS USED DURING **INSTRUMENT** M55,M112,T28,M81,BID00787, **PREPARATION** TID41094 Sample ID Tested 07/18/23 Tester **DTRA** Recorded 07/18/23 1.00837 1.00814 - 1.00859 1.00847 **Actual Density** GM/ML Sample ID Tested 07/18/23 Tester DTRA Recorded 07/18/23 3071474 Lot#

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager