

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9024			
SULFURIC ACID 0.04 NORMAL			
Test	Target/UOM	Range	Result
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		07/21/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.9
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		07/21/23
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.9
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
Actual Normality	0.0400 EQ/L	0.0395 - 0.0405	0.0395
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/21/23			
Tester JENC			
Recorded 07/21/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION		INSTRUMENT	GC, T222, M55, M91, BID00280, T ID41100
Sample ID			
Tested	07/21/23		
Tester	JENC		
Recorded	07/21/23		
EXPIRATION DATE		MM/DD/YY MM/DD/YYYY	07/30/25
Sample ID			
Tested	07/21/23		
Tester	JENC		
Recorded	07/21/23		
SHELF LIFE OF PRODUCT		SHELF LIFE TIME	2 YEARS
Sample ID			
Tested	07/21/23		
Tester	JENC		
Recorded	07/21/23		
Actual Density		0.99969 GM/ML	0.99949 - 0.99989 0.99980
Sample ID			
Tested	07/21/23		
Tester	JENC		
Recorded	07/21/23		
Lot #		3071953	
Made		07/21/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager