



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 1

Date: 07/24/23 at 11:46 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 1186557F BaraDemulsifier-IPA			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b> YES YES			
<b>FILE</b> YES			
Sample ID			
Tested 07/24/23			
Tester MMAT			
Recorded 07/24/23			
<b>Manufacturer's lot number Surfonc</b>		123456	E3121F35
<b>N-120</b>		LOT NUMBER	
Sample ID			
Tested 07/24/23			
Tester MMAT			
Recorded 07/24/23			
<b>Manufacture lot# Biosoft S-120</b>		123456	0008596138
		LOT NUMBER	
Sample ID			
Tested 07/24/23			
Tester MMAT			
Recorded 07/24/23			
<b>APPEARANCE AND COLOR</b>		CLEAR COLORLESS LIQUID	PASS
		APPEARANCE	
Sample ID			
Tested 07/24/23			
Tester MMAT			
Recorded 07/24/23			
<b>INSTRUMENTS USED DURING</b>		INSTRUMENT	M100,VF12-93
<b>PREPARATION</b>			
Sample ID			
Tested 07/24/23			
Tester MMAT			
Recorded 07/24/23			
<b>EXPIRATION DATE</b>		MM/DD/YYYY	07/30/25
		MM/DD/YYYY	
Sample ID			
Tested 07/24/23			
Tester MMAT			
Recorded 07/24/23			
<b>Lot #</b> 3072125			
<b>Made</b> 07/24/23			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 07/24/23 at 11:46 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**