Page: 1

Date: 07/24/23 at 11:46 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: HAC692F			
Calibration Standard 20 mg/L for EZ1022				
Test		Target/UOM	Range	Result
PREPARED TO FORM	IULATION ON	YES		YES
FILE		YES		
Sample II)			
Tested	07/24/23			
Tester	MMAT			
Recorded	07/24/23			
Calculated ppm Hydrogen Peroxide		20.00	19.50 - 20.50	20.12
		PPM		
Sample II)			
Tested	07/24/23			
Tester	MMAT			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
AI I LAIVANGE AND O	OLOIK	APPEARANCE		1 700
Sample II	2	AFFLAIVANCE		
Tested				
	07/24/23			
Tester	MMAT			
Recorded		INICTEL INVENIT		D70 \/F0 A44
INSTRUMENTS USED DURING		INSTRUMENT		P76,VF2-A11
PREPARATION				
Sample II				
Tested	07/24/23			
Tester	MMAT			
Recorded	07/24/23			
EXPIRATION DATE		MM/DD/YYYY		08/29/23
		MM/DD/YYYY		
Sample II)			
Tested	07/24/23			
Tester	MMAT			
Recorded	07/24/23			
Lot # 3072	2496			
	24/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I