



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS108F			
Conductivity Standard 149.6 umhos @25°C			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
CONDUCTIVITY 1 DEC	149.6	149.5 - 149.7	149.6
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
QCSP30: CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
	YES		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PAS
	APPEARANCE		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M92,T234
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
EXPIRATION DATE	MM/DD/YYYY		07/30/24
	MM/DD/YYYY		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
Lot #	3072554		
Made	07/25/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/25/23 at 2:25 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager