



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4350			
HYDROCHLORIC ACID 1.0 NORMAL N.I.S.T. TRACEABLE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
Actual Normality	1.0000 EQ/L	0.9995 - 1.0005	0.9996
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
+ Measurement Uncertainty	0.0079 EQ/L	= 0.0079	0.0079
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
Actual Density	1.01599 GM/ML	1.01513 - 1.01684	1.01608
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
N.I.S.T. Traceable to SRM 723 THAM	YES YES		YES
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			
CERTIFIED TOTAL ORGANIC CARBON (PPM)	1.0 PPM	0.0 - 3.0	2.2
Sample ID			
Tested 07/17/23			
Tester CCOA			
Recorded 07/17/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	07/17/23	
Tester	CCOA	
Recorded	07/17/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	T-72,GC,T-1000,M55,M44,BID0 0488
Sample ID		
Tested	07/17/23	
Tester	CCOA	
Recorded	07/17/23	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	07/30/25
Sample ID		
Tested	07/17/23	
Tester	CCOA	
Recorded	07/17/23	
SHELF LIFE OF PRODUCT	2 YEARS TIME	2 YEARS
Sample ID		
Tested	07/17/23	
Tester	CCOA	
Recorded	07/17/23	
COLOR	<10 PT-CO	6
Sample ID		
Tested	07/24/23	
Tester	CCOA	
Recorded	07/24/23	
Lot #	3071210	
Made	07/17/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager