

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS107F			
Conductivity Standard 74.0 umhos @25°C			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
CONDUCTIVITY @25 C.	74	73 - 75	74
	US/CM		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
QCSP30: CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
	YES		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
INSTRUMENTS USED DURING	INSTRUMENT		M92,T234
PREPARATION			
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
EXPIRATION DATE	MM/DD/YYYY		07/30/24
	MM/DD/YYYY		
Sample ID			
Tested 07/25/23			
Tester BPHI			
Recorded 07/25/23			
Lot #	3072553		
Made	07/25/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/25/23 at 1:45 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager