

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9765-4L			
WATER DEIONIZED REAGENT ACS GRADE/ASTM TYPE II (4L)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
MEETS ASTM TYPE II	YES		YES
SPECIFICATIONS	PASS		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
TOC < 50 PPB	< 50 PPB		<50 PPB
	PPB		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
SODIUM < 5 PPB	< 5 PPB		<5 PPB
	PPB		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
CHLORIDE < 5 PPB	< 5 PPB		<5 PPB
	PPB		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
SILICA < 3 PPB	< 3 PPB		<3 PPB
	PPB		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
HETEROTROPIC PLATE COUNT	ZERO		ZERO
ZERO	CFU		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			
CONDUCTIVITY	< 1 UMHOS		<1 UMHOS
	US/CM		
Sample ID			
Tested 07/27/23			
Tester TBUR			
Recorded 07/27/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

HEAVY METALS AS LEAD		< 10 PPB PPM	<10 PPB
Sample ID			
Tested	07/27/23		
Tester	TBUR		
Recorded	07/27/23		
PPM NO3		<0.4 PPM PPM	<0.4 PPM
Sample ID			
Tested	07/27/23		
Tester	TBUR		
Recorded	07/27/23		
PPM PHOSPHATE (PO4)		< 1.0 PPM PPM	<1.0 PPM
Sample ID			
Tested	07/27/23		
Tester	TBUR		
Recorded	07/27/23		
PPM SULFATE (SO4)		<1.0 PPM PPM	<1.0 PPM
Sample ID			
Tested	07/27/23		
Tester	TBUR		
Recorded	07/27/23		
EXPIRATION DATE		MM/DD/YY MM/DD/YYYY	07/30/24
Sample ID			
Tested	07/27/23		
Tester	TBUR		
Recorded	07/27/23		
SHELF LIFE OF PRODUCT		1 YEAR TIME	1 YEAR
Sample ID			
Tested	07/27/23		
Tester	TBUR		
Recorded	07/27/23		
Lot #		3071700	
Made		07/20/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 08/09/23 at 12:01 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager