



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> 9764-10L			
	WATER, DEIONIZED REAGENT GRADE ASTM TYPE III (10LC)		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>PREPARED TO ASTM D 1193</b>	PASS		PASS
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>CONDUCTIVITY</b>	<= 0.25 UMHOS US/CM		<=0.25 UMHOS
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>RESISTIVITY 18 MEG-OHM AT TIME OF PACKAG</b>	>= 4.0 MEGOHM PASS		>= 4.0 MEGOHM
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>TOC &lt; 200 ppb</b>	<= 200 PPB < 200 PPB		<= 200 PPB
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>Sodium (Na) &lt;=10 ppb</b>	<= 10 PPB <= 10 PPB		<= 10 PPB
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>Chloride (Cl) &lt;= 10 ppb</b>	<= 10 PPB <= 10 PPB		<= 10 PPB
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			
<b>Silica &lt;= 500 ppb</b>	<= 500 PPB <= 500 PPB		<= 500 PPB
Sample ID			
Tested 08/03/23			
Tester TBUR			
Recorded 08/03/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>MEETS ASTM D 1193 TYPE III SPECIFICATION</b>	PASS PASS	PASS
Sample ID		
Tested	08/03/23	
Tester	TBUR	
Recorded	08/03/23	
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR TIME	1 YEAR
Sample ID		
Tested	08/03/23	
Tester	TBUR	
Recorded	08/03/23	
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	07/30/24
Sample ID		
Tested	08/03/23	
Tester	TBUR	
Recorded	08/03/23	
<b>Lot #</b>	30630113	
<b>Made</b>	07/10/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**