



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7435			
POTASSIUM PERMANGANATE 0.05 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0500 EQ/L	0.0495 - 0.0505	0.0500
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
QCSP9:STADARDIZATION OF POTASSIUM PERMAN	YES QCSP9		YES
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
N.I.S.T. Traceable to SRM 40	YES YES		YES
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
APPEARANCE AND COLOR	CLEAR PURPLE LIQUID APPEARANCE		PASS
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,VF4-N29
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
Actual Density	0.99934 GM/ML	0.99910 - 0.99959	0.99951
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/24
Sample ID			
Tested	07/31/23		
Tester	JENC		
Recorded	07/31/23		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

1 YEAR
 TIME

1 YEAR

Sample ID

Tested 07/31/23

Tester JENC

Recorded 07/31/23

Lot # 3072744

Made 07/31/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager