

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7245			
POTASSIUM HYDROXIDE 0.1 NORMAL (MOLAR) IN IPA (ASTM D 974)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/03/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	23.2
SAMPLING DEGC	DEG C		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/03/23
	MM/DD/YYYY		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	23.2
DEG C	DEG C		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
PREPARED TO ASTM D 974	YES		YES
	YES		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			
QCSP7: STANDARDIZATION OF	YES		YES
AQUEOUS BASES	METHOD		
Sample ID			
Tested 08/03/23			
Tester JENC			
Recorded 08/03/23			


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

NORMALITY TO 4 DEC PLACES	0.1000	0.0995 - 0.1005	0.1005
	EQUIV/LITER		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
Actual Density	0.79116	0.79031 - 0.79201	0.79161
	GM/ML		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE TO SLIGHT TURBID APPEARANCE		PASS
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M91,BID02259,TID41100,M91, M103,T77	
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/24
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
Lot #	3073123		
Made	08/03/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager