

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8410 SILVER NITRATE 0.0282 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/03/23			
Tester CCOA			
Recorded 08/03/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/03/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	21.9
SAMPLING DEGC	DEG C		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/03/23
	MM/DD/YYYY		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.3
DEG C	DEG C		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			
Actual Normality	0.0282	0.0279 - 0.0285	0.0282
	EQ/L		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			
QCSP54:STANDARDIZATION OF	YES		YES
LOW NORMAL SN	QCSP54		
Sample ID			
Tested 08/03/23			
Tester MMAT			
Recorded 08/03/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 919 NaCl

YES

YES

YES

Sample ID

Tested 08/03/23

Tester MMAT

Recorded 08/03/23

Actual Density

1.00225

1.00203 - 1.00247

1.00232

GM/ML

Sample ID

Tested 08/03/23

Tester MMAT

Recorded 08/03/23

APPEARANCE AND COLOR

CLEAR WATER WHITE

PASS

LIQUID

APPEARANCE

Sample ID

Tested 08/03/23

Tester MMAT

Recorded 08/03/23

**INSTRUMENTS USED DURING
PREPARATION**

INSTRUMENT

M55,M108,VF5-E3

Sample ID

Tested 08/03/23

Tester MMAT

Recorded 08/03/23

EXPIRATION DATE

MM/DD/YYYY

08/30/25

MM/DD/YYYY

Sample ID

Tested 08/03/23

Tester MMAT

Recorded 08/03/23

Lot # 3080114

Made 08/03/23

Quality Manager