



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 9026 SULFURIC ACID 0.03 NORMAL			
Test	Target/UOM	Range	Result
Prepared to formulation on File	Yes YES		YES
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		08/02/23
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	22.6
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		08/02/23
	MM/DD/YYYY		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	22.6
<b>DEG C</b>	DEG C		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>Actual Normality</b>	0.0300	0.0295 - 0.0305	0.0302
	EQ/L		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			
<b>N.I.S.T. Traceable to SRM 723 THAM</b>	YES		YES
	YES		
Sample ID			
Tested 08/02/23			
Tester MHUS			
Recorded 08/02/23			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 08/02/23 at 2:39 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**QCSP5: STANDARDIZATION OF  
AQUEOUS ACIDS**

YES  
METHOD

YES

Sample ID  
Tested 08/02/23  
Tester MHUS  
Recorded 08/02/23

**APPEARANCE AND COLOR**

CLEAR WATER WHITE  
LIQUID  
APPEARANCE

PASS

Sample ID  
Tested 08/02/23  
Tester MHUS  
Recorded 08/02/23

**INSTRUMENTS USED DURING  
PREPARATION**

INSTRUMENT

T-42,M55,BID04099,PIDM91

Sample ID  
Tested 08/02/23  
Tester MHUS  
Recorded 08/02/23

**Lot #** 3080135

**Made** 08/02/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**