Page: 1

Date: 08/03/23 at 5:50 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	Product:	ND644F			
			ogen Standard 10.0 ppm		
Test			Target/UOM	Range	Result
PREPARED	TO FORMUL	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	08/03/23			
	Tester	MMAT			
	Recorded	08/03/23			
PPM NITRO	GEN		10.00	9.90 - 10.10	10.00
			PPM		
	Sample ID				
	Tested	08/03/23			
	Tester	MMAT			
	Recorded	08/03/23			
APPEARANG			CLEAR LIQUID		PASS
A I LAWA	)		APPEARANCE		17100
	Sample ID		ALL EXIVANCE		
		00/00/00			
	Tested	08/03/23			
	Tester	MMAT			
	Recorded	08/03/23	NOTELINATIA		D77 \/E500 \/04
INSTRUMENTS USED DURING			INSTRUMENT		P77,VF500-X31
PREPARATION					
	Sample ID				
	Tested	08/03/23			
	Tester	MMAT			
	Recorded	08/03/23			
<b>EXPIRATION</b>	I DATE		MM/DD/YYYY		11/03/23
			MM/DD/YYYY		
	Sample ID				
	Tested	08/03/23			
	Tester	MMAT			
	Recorded	08/03/23			
Lot	<b>#</b> 30803				
Made	08/03	/ 43			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

lauland I