

Page: 1

Date: 08/01/23 at 1:50 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Prod	uct: 1760			
	CERIC	SULFATE TITRANT 0.1 NORM		
Test		Target/UOM	Range	Result
PREPARED TO FO	RMULATION ON	YES		YES
FILE		YES		
Sample	e ID			
Tested	08/01/23			
Tester	JENC			
Record	ded 08/01/23			
Actual Normality		0.1000	0.0995 - 0.1005	0.1001
		EQ/L		
Sampl	e ID			
Tested				
Tester				
Record				
Actual Density		1.08960	1.08631 - 1.09289	1.08980
		GM/ML		
Sampl	a ID	S,		
Tested				
Tester	JENC			
Record				
		YES		YES
QCSP19:STANDARDIZATION OF CERIC SULFATE		QCSP19		123
	. ID	QCSF 19		
Sample				
Tested				
Tester				
Record		OLEAR ORANGE LIQUIR		DA 00
APPEARANCE AND COLOR		CLEAR ORANGE LIQUID APPEARANCE		PASS
Sampl	e ID			
Tested				
Tester				
Record				
INSTRUMENTS US		INSTRUMENT		M112,VF5-L13,M55,M91,BID02
PREPARATION				259,TID4
Sample	a ID			200,1121
Tested				
Tester				
r ester Record				
EXPIRATION DATE		MM/DD/YY		02/28/24
EAFIRATION DATE		MM/DD/YYYY		02/20/24
0 1	- ID	IVIIVI/UU/I I I I I		
Sample				
Tested				
Tester				
Record				
	72875			
Made 08	/01/23			

Page: 2

Date: 08/01/23 at 1:50 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager