

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: S2675			
SODIUM CHLORIDE REAGENT ACS CRYSTAL			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	$\geq 99.0\%$ %		100.00%
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
PH OF A 5% SOLUTION	5.0 - 9.0 AT 25 DEG C PH		5.9
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
INSOLUBLE MATTER	MAX 0.005% %		0.0010%
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
IODIDE	MAX 0.002% PPM		$\leq 0.0020\%$
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
BROMIDE (BR)	MAX 0.01% PPM		$< 0.010\%$
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
PHOSPHATE CONTENT	MAX 5 PPM PPM		< 5 PPM
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
SULFATE (SO4) CONTENT	MAX 0.004% MAX 60 MG/KG		$< 0.0040\%$
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			
BARIUM	PASSES TEST BARIUM		PASSES TEST
Sample ID			
Tested 07/31/23			
Tester TBUR			
Recorded 07/31/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/31/23 at 10:05 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

HEAVY METAL (AS PB)

MAX 5 PPM
PPM

<5.0 PPM

Sample ID

Tested 07/31/23

Tester TBUR

Recorded 07/31/23

IRON

MAX 2 PPM
PPM

<2 PPM

Sample ID

Tested 07/31/23

Tester TBUR

Recorded 07/31/23

CALCIUM CONTENT

MAX 0.002%
PPM

<0.0010%

Sample ID

Tested 07/31/23

Tester TBUR

Recorded 07/31/23

MAGNESIUM (MG) CONTENT

MAX 0.001%
PPM

<0.0010%

Sample ID

Tested 07/31/23

Tester TBUR

Recorded 07/31/23

POTASSIUM CONTENT

MAX 0.005%
PPM

0.0030%

Sample ID

Tested 07/31/23

Tester TBUR

Recorded 07/31/23

DATE OF MANUFACTURE

DD/MM/YYYY

03/24/23

Sample ID

Tested 07/31/23

Tester TBUR

Recorded 07/31/23

Lot # 320902

Made 07/28/23

Quality Manager