

Page: 1

Date: 08/04/23 at 4:00 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 8650 SODIUM HYDROXIDE 1.0 NORMAL NIST TRACEABLE SOLUTION				
Test	Target/UOM	Range	Result	
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	26.0	
SAMPLING DEGC	DEG C			
Sample ID				
Tested 08/04/23				
Tester BPHI				
Recorded 08/04/23				
Actual Normality	1.0000	0.9995 - 1.0005	1.0000	
•	EQ/L			
Sample ID				
Tested 08/04/23				
Tester BPHI				
Recorded 08/04/23				
+- Measurement Uncertainty	0.0070	= 0.0070	0.0070	
• • • • •	EQ/L			
Sample ID				
Tested 08/04/23				
Tester BPHI				
Recorded 08/04/23				
CSP7: STANDARDIZATION OF	YES		YES	
AQUEOUS BASES	METHOD			
Sample ID				
Tested 08/04/23				
Tester BPHI				
Recorded 08/04/23				
VII.S.T. Traceable to SRM 84	Yes		YES	
VII.O.T. Traddable to Ortin 0-4	YES		120	
Sample ID	120			
Tested 08/04/23				
Tester BPHI				
Recorded 08/04/23				
Actual Density	1.04114	1.04049 - 1.04179	1.04119	
iotaai Dolloity	GM/ML	1.0-00-0 - 1.0-1/18	1.04113	
Sample ID	OIVI/IVIL			
Tested 08/04/23 Tester BPHI				
Recorded 08/04/23				
PREPARED TO FORMULATION ON	YES		YES	
FILE	YES		ILO	
	125			
Sample ID Tested 08/04/23				
Tested 08/04/23 Tester BPHI				
Recorded 08/04/23				
SAMPLE COLLECTION & HANDLING	YES		YES	
	YES		IES	
PER LAB040	150			
Sample ID				
Tested 08/04/23				
Tester BPHI				
Recorded 08/04/23				



Made 08/04/23

Page: 2

Date: 08/04/23 at 4:00 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

		Customer PO # :	
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/04/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
AMBIENT TEMPERATURE DURING	3 25.0	10.0 - 35.0	26.9
SAMPLING DEGC	DEG C		
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/04/23
	MM/DD/YYYY		33/0 1/20
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	21.5
DEG C	DEG C		
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
INSTRUMENTS USED DURING	INSTRUMENT		M55,M25,T-IV,M91,BID00535,T
PREPARATION			ID41111
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
EXPIRATION DATE	MM/DD/YY		08/30/25
	MM/DD/YYYY		33/33/23
Sample ID	18.18.10 D) 1 1 1 1		
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
SHELF LIFE OF PRODUCT	2 YEARS		2 YEARS
	TIME		ZILMO
Sample ID			
Tested 08/04/23			
Tester BPHI			
Recorded 08/04/23			
Lot # 3080269			
00/01/00			



Page: 3

Date: 08/04/23 at 4:00 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager