

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3550			
EDTA TITRANT 0.01 MOLAR (0.02 NORMAL) (1ML = 1MG CaCO3) NIST			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
Actual Molarity	0.0100	0.0095 - 0.0105	0.0098
	MOL/L		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES		YES
	QCSP4		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
Actual Density	1.00009	0.99990 - 1.00028	1.00028
	GM/ML		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
N.I.S.T. Traceable to SRM 915	YES		YES
	YES		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,VF12-96,M55,M91,BID022 59,TID41
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			
EXPIRATION DATE	MM/DD/YY		08/30/25
	MM/DD/YYYY		
Sample ID			
Tested 07/31/23			
Tester JENC			
Recorded 07/31/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/31/23 at 10:18 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

24 MONTHS
TIME

2 YEARS

Sample ID

Tested 07/31/23

Tester JENC

Recorded 07/31/23

Lot # 3072855

Made 07/31/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager