

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged F	Product:	SPE510F				
-		SIMULATED	INTESTINAL FLUID USP			
Test			Target/UOM	Range	Result	
PREPARED TO	J FORMUL	ATION ON	YES		YES	
FILE			YES			
S	ample ID					
T	ested	07/31/23				
Т	ester	RSHE				
R	ecorded	07/31/23				
PH OF PRODU	JCT/SOLU	ΓΙΟΝ	6.800	6.700 - 6.900	6.804	
			PH			
S	ample ID					
Т	ested	07/31/23				
Т	ester	RSHE				
R	ecorded	07/31/23				
INSTRUMENTS USED DURING			INSTRUMENT		M103,GC,EF1-PA	
PREPARATION						
S	ample ID					
T	ested	07/31/23				
T	ester	RSHE				
R	ecorded	07/31/23				
Lot #	Lot # 3072821					
Made	07/31,					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stol

Quality Manager