

Page: 1

Date: 08/04/23 at 4:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

D. I. D. I. 062		- π ·			
Packaged Product: 8628 SODIUM HYDROXIDE 0.1 NORMAL IN METHANOL					
Test	Target/UOM	Range	Result		
Actual Normality	0.1000	0.0995 - 0.1005	0.0998		
Actual Normality	EQ/L	0.0000 0.1000	0.0000		
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
QCSP7: STANDARDIZATION OF	YES		YES		
AQUEOUS BASES	METHOD				
Sample ID	2				
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
N.I.S.T. Traceable to SRM 84 (KH			YES		
Talion I Haceable to Shiw 64 (Kh	YES		113		
Comple ID	163				
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23			VEO		
PREPARED TO FORMULATION C			YES		
FILE	YES				
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
SAMPLE COLLECTION & HANDL			YES		
PER LAB040	YES				
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/04/23		
PURPOSES	MM/DD/YYYY				
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
AMBIENT TEMPERATURE DURIN	IG 25.0	10.0 - 35.0	23.0		
SAMPLING DEGC	DEG C				
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/04/23		
	MM/DD/YYYY				
Sample ID					
Tested 08/04/23					
Tester CCOA					
Recorded 08/04/23					
1 tecorded 00/04/23					

08/04/23

Made

Page: 2

Date: 08/04/23 at 4:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

AMBIENT TEST TEMPER	RATURE	20.0 DEG C	15.0 - 30.0	23.2
Sample ID		= = 5. 5		
Tested	08/04/23			
Tester	CCOA			
Recorded	08/04/23			
APPEARANCE AND COL	_OR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	08/04/23			
Tester	CCOA			
Recorded	08/04/23			
INSTRUMENTS USED D	URING	INSTRUMENT		M112,GC,VF5-E5
PREPARATION				
Sample ID				
Tested	08/04/23			
Tester	CCOA			
Recorded	08/04/23			
Actual Density		0.81669	0.81600 - 0.81727	0.81622
		GM/ML		
Sample ID				
Tested	08/04/23			
Tester	CCOA			
Recorded	08/04/23			
EXPIRATION DATE		MM/DD/YYYY		08/30/24
		MM/DD/YYYY		
Sample ID				
Tested	08/04/23			
Tester	CCOA			
Recorded	08/04/23			
Lot # 30802	68			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stop

Quality Manager