Page: 1

Date: 08/01/23 at 1:04 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: LY125F			
		STANDARD 1.0 PPM AQUE	OUS	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample II)			
Tested	08/01/23			
Tester	MMAT			
Recorded	08/01/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample II)			
Tested	08/01/23			
Tester	MMAT			
Recorded	08/01/23			
INSTRUMENTS USED	DURING	INSTRUMENT		P124,VF100-K11
PREPARATION				
Sample II)			
Tested	08/01/23			
Tester	MMAT			
Recorded				
Actual ppm Cyanide		1.00	0.98 - 1.02	1.01
,		PPM		
Sample II)			
Tested	08/01/23			
Tester	MMAT			
Recorded				
EXPIRATION DATE	00/01/20	MM/DD/YY		10/06/23
ZXI IVANION DAIL		MM/DD/YYYY		10/00/20
Sample II	1	, 55, 1111		
Tested	08/01/23			
Tester	MMAT			
Recorded				
Lot # 3080				
Made 08/0	1/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I