



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 08/09/23 at 6:07 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY386F SULFUR STANDARD 10.0 PPM W/W IN ISOCTANE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/09/23			
Tester MMAT			
Recorded 08/09/23			
Calculated ppm Sulfur	10.00 PPM	9.80 - 10.20	10.00
Sample ID			
Tested 08/09/23			
Tester MMAT			
Recorded 08/09/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/09/23			
Tester MMAT			
Recorded 08/09/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P113,VF1-Z2
Sample ID			
Tested 08/09/23			
Tester MMAT			
Recorded 08/09/23			
Lot #	3080803		
Made	08/09/23		

Quality Manager