

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

		SODIUM SULFITE 1.0 MOLAR PH 10.42 ± 0.02			
Test			Target/UOM	Range	Result
	D TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	08/10/23			
	Tester	RSHE			
	Recorded	08/10/23			
PH OF PRODUCT/SOLUTION		JTION	10.420	10.400 - 10.440	10.424
			PH		
	Sample ID				
	Tested	08/10/23			
	Tester	RSHE			
	Recorded	08/10/23			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	08/10/23			
	Tester	RSHE			
	Recorded	08/10/23			
INSTRUMENTS USED DURING		URING	INSTRUMENT		M103,T-199
REPARA	ΓΙΟΝ				
	Sample ID				
	Tested	08/10/23			
	Tester	RSHE			
	Recorded	08/10/23			
Lo	t# 30808	83			
Ma		/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stot

Quality Manager