

Page: 1

Date: 08/09/23 at 6:04 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	6909			
	CONDUC'	TIVITY STANDARD 1,250 uMH		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23	40-0	4050	4070
CONDUCTIVITY @25 C.		1250	= 1250	1250
		US/CM		
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23	V=0		V=0
QCSP30:CONDUCTIVIT	Y	YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23	V=0		\/ <u></u>
TRACEABLE TO N.I.S.T	. SRM 3191	YES		YES
		YES		
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23			
APPEARANCE AND CO	LOR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23			
INSTRUMENTS USED D	URING	INSTRUMENT		M92,T234
PREPARATION				
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23			
EXPIRATION DATE		MM/DD/YYYY		08/30/24
		MM/DD/YYYY		
Sample ID				
Tested	08/09/23			
Tester	BPHI			
Recorded	08/09/23			
Lot # 30808				
Made 08/09	9/23			

Page: 2

Date: 08/09/23 at 6:04 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager