



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS120F Mixed Standard in Acetonitrile			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
W/W% DIISOBUTYL PHTHALATE	8.0000 W/W%	7.9500 - 8.0500	8.0000
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
W/W% DIETHYL PHTHALATE	0.6000 W/W%	0.5900 - 0.6100	0.6000
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
W/W PHTHALIC ANHYDRIDE	0.4000 W/W%	0.3800 - 0.4200	0.4000
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
W/W ETHYL BENZOATE	7.0000 W/W%	6.9000 - 7.1000	7.0000
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
APPEARANCE AND COLOR	CLEAR TO PALE YELLOW LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M115,EF2-S8
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		08/30/25
Sample ID			
Tested 08/11/23			
Tester MMAT			
Recorded 08/11/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/11/23 at 9:03 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	3080438
Made	08/11/23

Quality Manager