

Made 08/10/23

Page: 1

Date: 08/10/23 at 6:08 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Customer PO # :					
Packag	ed Product:	M-528F			
		AMMONIA	NITROGEN STANDARD 25 F	PPM AS NH3	
Test			Target/UOM	Range	Result
PREPARE	ED TO FORMULA	TION ON	YES	_	YES
FILE			YES		
	Sample ID				
	Tested 08	3/10/23			
1	Tester M	MAT			
	Recorded 08	3/10/23			
CERTIFIED PPM NH3			25.0	24.0 - 26.0	25.5
			PPM		
	Sample ID				
	Tested 08	3/10/23			
	Tester M	MAT			
	Recorded 08	3/10/23			
APPEARANCE AND COLOR		R	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested 08	3/10/23			
	Tester M	MAT			
	Recorded 08	3/10/23			
INSTRUMENTS USED DURING		RING	INSTRUMENT		M115,VF3-J19
PREPARA	ATION				
	Sample ID				
	Tested 08	3/10/23			
	Tester M	MAT			
	Recorded 08	3/10/23			
SHELF LIFE OF PRODUCT		3 MONTHS		3 MONTTHS	
			TIME		
	Sample ID				
		3/10/23			
	Tester M	MAT			
	Recorded 08	3/10/23			
EXPIRATION DATE			MM/DD/YYYY		11/30/23
			MM/DD/YYYY		
	Sample ID				
	•	3/10/23			
	Tester M	MAT			
	Recorded 08	3/10/23			
Le	ot# 3081068	}			



Page: 2

Date: 08/10/23 at 6:08 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager