

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged I	Product:	UCT126F			
			STD FOR TSS 40 MG/L SOLU	FION	
Test		-	Target/UOM	Range	Result
PREPARED T	O FORMUL	ATION ON	YES		YES
FILE			YES		
S	Sample ID				
Т	ested	08/07/23			
Т	ester	AROS			
F	Recorded	08/07/23			
ACTUAL PPM			40.00	39.50 - 40.50	40.10
			PPM		
S	ample ID				
Т	ested	08/07/23			
Т	ester	AROS			
F	Recorded	08/07/23			
APPEARANCE AND COLOR		.OR	SLIGHTLY TURBID LIQUID		PASS
			APPEARANCE		
S	Sample ID				
Т	ested	08/07/23			
Т	ester	AROS			
F	Recorded	08/07/23			
INSTRUMENTS USED DURING		URING	INSTRUMENT		M115,VF4-N21
PREPARATIO	N				
S	Sample ID				
Т	ested	08/07/23			
Т	ester	AROS			
F	Recorded	08/07/23			
EXPIRATION DATE			MM/DD/YY		02/28/24
			MM/DD/YYYY		
S	ample ID				
Т	ested	08/07/23			
Т	ester	AROS			
		08/07/23			
Lot #	30807	13			
Made	08/07				
made					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA