

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 1520			
BUFFER REFERENCE pH 9.0			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>PH OF PRODUCT/SOLUTION</b>	9.000	8.980 - 9.020	8.995
	PH		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>QCSP 20 PH CALIBRATION</b>	YES		YES
<b>METHODS</b>	METHOD		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>Actual Temp</b>	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>N.I.S.T. Traceable</b>	YES		YES
	YES		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>NIST TRACEABLE TO SRM186I AND 186II</b>	YES		YES
	YES		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>PH NIST TRACEABLE TO SRM 187</b>	YES		YES
	YES		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			
<b>TRACEABLE TO NIST SRM 191I AND SRM 191II</b>	YES		YES
	YES		
Sample ID			
Tested 08/10/23			
Tester LWIL			
Recorded 08/10/23			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 08/10/23 at 11:42 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>APPEARANCE AND COLOR</b>	CLEAR TO PALE YELLOW LIQUID APPEARANCE	PASS
Sample ID		
Tested	08/10/23	
Tester	LWIL	
Recorded	08/10/23	
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M103,M112,T-245
Sample ID		
Tested	08/10/23	
Tester	LWIL	
Recorded	08/10/23	
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	08/30/24
Sample ID		
Tested	08/10/23	
Tester	LWIL	
Recorded	08/10/23	
<b>SHELF LIFE OF PRODUCT</b>	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested	08/10/23	
Tester	LWIL	
Recorded	08/10/23	
<b>Lot #</b>	3080230	
<b>Made</b>	08/10/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**