

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: 0045		esserble Colution	
Test	c Acid 2.0 Normal NIST Tr Target/UOM	Range	Result
Actual Normality	2.0000	1.9990 - 2.0010	2.0004
,	EQ/L		
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
REPARED TO FORMULATION ON	YES		YES
ILE	YES		
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
CSP37:STANDARDIZATION	YES		YES
CETIC ACID SOLN	QCSP37		120
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
PPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
FFEARANCE AND COLOR	LIQUID		FASS
	APPEARANCE		
	APPEARANCE		
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
ISTRUMENTS USED DURING	INSTRUMENT		M112,VF3-J20
REPARATION			
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
ctual Density	1.01404	1.00529 - 1.02278	1.01513
	GM/ML		
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
XPIRATION DATE	MM/DD/YYYY		02/28/25
	MM/DD/YYYY		
Sample ID			
Tested 08/08/23			
Tester CCOA			
Recorded 08/08/23			
Lot # 3080757			
Made 08/08/23			



Page: 2 Date: 08/08/23 at 3:09 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager