

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

	CERTC	SULFATE 0.2 NORMAL IN 2	NORMAL SOLLORIC ACID	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
Actual Normality		0.2000	0.1985 - 0.2015	0.1994
		EQ/L		
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
Actual Density		1.13095	1.12805 - 1.13385	1.13014
		GM/ML		
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
QCSP19:STANDARDIZ	ATION OF	YES		YES
CERIC SULFATE		QCSP19		
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
APPEARANCE AND CO	LOR	CLEAR ORANGE LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
INSTRUMENTS USED D	URING	INSTRUMENT		M55,M87,VF2-B18
PREPARATION				
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
EXPIRATION DATE		MM/DD/YY MM/DD/YYYY		02/28/24
Sample ID				
Tested	08/14/23			
Tester	LWIL			
Recorded	08/14/23			
Lot # 30808	329			
Made 08/14				



Page: 2 Date: 08/14/23 at 11:32 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager