

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4133 HYDROCHLORIC ACID 0.1 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/16/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	22.2
SAMPLING DEGC	DEG C		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/16/23
	MM/DD/YYYY		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.2
DEG C	DEG C		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
Actual Normality	0.1000	0.0991 - 0.1009	0.0996
	EQ/L		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			
QCSP5: STANDARDIZATION OF	YES		YES
AQUEOUS ACIDS	METHOD		
Sample ID			
Tested 08/16/23			
Tester MHUS			
Recorded 08/16/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 723 THAM

YES

YES

YES

Sample ID

Tested 08/16/23

Tester MHUS

Recorded 08/16/23

APPEARANCE AND COLOR

CLEAR LIQUID

PASS

APPEARANCE

Sample ID

Tested 08/16/23

Tester MHUS

Recorded 08/16/23

INSTRUMENTS USED DURING PREPARATION

INSTRUMENT

M115,VF2-B18

Sample ID

Tested 08/16/23

Tester MHUS

Recorded 08/16/23

EXPIRATION DATE

MM/DD/YYYY

11/30/23

MM/DD/YYYY

Sample ID

Tested 08/16/23

Tester MHUS

Recorded 08/16/23


Lot # 3081104

Made 08/16/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager