

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	CY198F			
	AMMONIUM	NITROGEN STANDARD 5 PI	PM AS NH3	
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested (08/15/23			
Tester A	AROS			
Recorded (08/15/23			
CERTIFIED PPM NH3		5.0	4.0 - 6.0	4.8
		PPM		
Sample ID				
	08/15/23			
	AROS			
	08/15/23			
APPEARANCE AND COLO		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
	08/15/23			
	AROS			
	08/15/23			
INSTRUMENTS USED DU		INSTRUMENT		P83,VF1-Z33
PREPARATION				100,011200
Sample ID Tested (
	08/15/23 AROS			
	08/15/23 -	6 MONTHS		6 MONTHS
SHELF LIFE OF PRODUC	1			0 MONTHS
- · ·-		TIME		
Sample ID				
	08/15/23			
	AROS			
	08/15/23			
EXPIRATION DATE		MM/DD/YYYY		02/28/24
		MM/DD/YYYY		
Sample ID				
	08/15/23			
	AROS			
Recorded (08/15/23			
Lot # 308156	5			
Made 08/15/				
	20			



Page: 2 Date: 08/15/23 at 4:43 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager