

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	3858			
FLUORIDE STANDARD 1.0 PPM				
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION	YES	-	YES
		YES		
Sample ID				
Tested	08/18/23			
Tester	CCOA			
Recorded	08/18/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	08/18/23			
Tester	CCOA			
Recorded	08/18/23			
INSTRUMENTS USED D	URING	INSTRUMENT		P-76,VF2-B10,M79
PREPARATION				
Sample ID				
Tested	08/18/23			
Tester	CCOA			
Recorded	08/18/23			
PPM FLUORIDE		1.00	0.85 - 1.15	1.07
		PPM		
Sample ID				
Tested	08/18/23			
Tester	CCOA			
Recorded	08/18/23			
EXPIRATION DATE	00/10/20	MM/DD/YY		02/28/24
		MM/DD/YYYY		02,20,24
Sample ID				
Tested	08/18/23			
	08/18/23 CCOA			
Tester Recorded	08/18/23			
SHELF LIFE OF PRODU		6 MONTHS		6 MONTHS
		TIME		6 MONTHS
Sample ID				
Tested	08/18/23			
Tester	CCOA			
Recorded	08/18/23			
Lot # 30817	10			
Made 08/18				
	, 20			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard Stop

Quality Manager